

# HANNA instruments

HI 7031 Conductivity Calibration Solution 1413  $\mu\text{s}/\text{cm}$

## Health & Safety data sheets

### SECTION 1 - IDENTIFICATION OF PREPARATION

**Product name:**

- 1,413  $\mu\text{s}$  Standard solution

**Application:**

- Conductivity calibration solution

**Cat. No. :**

- HI7031

**Manufacturer  
identification:**

HANNA INSTRUMENTS  
V.le delle Industrie 12  
Ronchi di Villafranca  
PADOVA ITALY

### SECTION 2 - COMPOSITION/INFORMATION ON INGREDIENTS

**Chemical characterization:**

- Solution in water

**Chemical composition:**

- Potassium chloride: < 5.0%

### SECTION 3 - HAZARD IDENTIFICATION

N/A

### SECTION 4- FIRST AID MEASURES

**Eye contact:**

- Irrigate thoroughly with water. If discomfort persists obtain medical attention.

**Skin contact:**

- Wash off thoroughly with soap and water.

**Ingestion:**

- Wash out mouth thoroughly with water. In severe cases obtain medical attention.

**Inhalation:**

- N/A

THE INFORMATION CONTAINED HEREIN IS BASED ON DATA CONSIDERED TO BE ACCURATE. HOWEVER, NO WARRANTY IS EXPRESSED OR IMPLIED REGARDING THE ACCURACY OF THESE DATA OR THE RESULTS TO BE OBTAINED FROM THE USE THEREOF.

# HANNA instruments

HI 7031 Conductivity Calibration Solution 1413  $\mu\text{s}/\text{cm}$

## Health & Safety data sheets

### SECTION 5 - FIRE FIGHTING MEASURES

**Special risks:**

- N/A

**Fire extinguisher:**

- N/A

**Special fire fighting precautions:**

- None

### SECTION 6 - ACCIDENTAL RELEASE MEASURES

**In case of accidental spill or release:**

- Wear appropriate protective clothing
- If local regulations permit, mop up with plenty of water and run to waste, diluting greatly with running water. Otherwise transfer to container and arrange removal by disposal company. Wash site of spillage thoroughly with detergent and water.

### SECTION 7- HANDLING AND STORAGE

**Handling:**

- Under no circumstances eat or drink while handling this material
- Wash hands and face after working with material

**Storage:**

- Store at room temperature (@ 15-25°C)
- Protect from direct sunlight and moisture
- Keep container well closed
- Keep away from children

### SECTION 8- EXPOSURE CONTROL/PERSONAL PROTECTION

As appropriate to quantity handled

**Respiratory protection:**

- Not applicable

**Protective gloves:**

- Rubber or plastic

**Eye protection:**

- Goggles/face shield

### SECTION 9- PHYSICAL/CHEMICAL PROPERTIES

**Appearance and odor:**

- Clear and colorless liquid
- Odorless

**Water reactivity:**

- Non-reactive

**Solubility in water:**

- Miscible in any proportion

**Melting point:**

- N/A

**Boiling point:**

- > 100 C°

THE INFORMATION CONTAINED HEREIN IS BASED ON DATA CONSIDERED TO BE ACCURATE. HOWEVER, NO WARRANTY IS EXPRESSED OR IMPLIED REGARDING THE ACCURACY OF THESE DATA OR THE RESULTS TO BE OBTAINED FROM THE USE THEREOF.

# **HANNA instruments**

**HI 7031 Conductivity Calibration Solution 1413 µs/cm**

## **Health & Safety data sheets**

### SECTION 10- STABILITY AND REACTIVITY

**Stability:**

- Stable

### SECTION 11- TOXICOLOGICAL INFORMATION

**In case of ingestion:**

- After ingestion of large amounts: nausea, vomiting, cardiac dysrhythmia, cardiovascular failure. However, when handled properly, hazardous effects are unlikely to occur.

- **Further data:**

LD50 (oral-rat) N/A

NO evidence of carcinogenic properties. No evidence of mutagenic or teratogenic effects.

### SECTION 12- ECOLOGICAL INFORMATION

No environmental hazard is anticipated provided that the material is handled and disposed of with due care and attention

### SECTION 13- DISPOSAL CONSIDERATIONS

**Waste disposal:**

- Chemical residues are generally classified as special waste and thus covered by local regulations. Contact authorities or disposal companies for advice.
- In compliance with current regulations.

### SECTION 14- TRANSPORT INFORMATION

UN-No.:

IMDG class: NR

IMO: NR

Packaging group:

IATA: NR

Packaging group:

Correct technical name: Not-Restricted

ADR/RID: NR

THE INFORMATION CONTAINED HEREIN IS BASED ON DATA CONSIDERED TO BE ACCURATE. HOWEVER, NO WARRANTY IS EXPRESSED OR IMPLIED REGARDING THE ACCURACY OF THESE DATA OR THE RESULTS TO BE OBTAINED FROM THE USE THEREOF.

# **HANNA instruments**

**HI 7031 Conductivity Calibration Solution 1413  $\mu\text{s}/\text{cm}$**

## **Health & Safety data sheets**

### SECTION 15- REGULATORY INFORMATION

#### **Labeling according to EEC directives:**

R-phrases: -

S-phrases: -

### SECTION 16- OTHER INFORMATION

Emergency telephone number 0039-49-9070211

last rev. June 2002